



DESNA CONTROLS & WEIGHING INC.
 412 MANITOU ROAD SE, CALGARY, AB T2G 4C4
 Phone: 403/243-4848 Fax: 403/243-9338

SAMPLE

SCALE INSPECTION REPORT:

CUSTOMER: Company Name

DATE: 6-Jan-08

ADDRESS: 1230 Macleod Trail
 Calgary, AB

TECHNICIAN: G. Brown

This is to certify that the following weighing equipment has been verified/calibrated by Desna Controls & Weighing Inc. and the device has been found to provide a readout accurate to within +/- 0.1% of the indicated capacity unless otherwise noted.*
 This system was checked using government approved test weights with verification traceable to the NIST in accordance with ASTM E4-99

INDICATOR: IQ-355


SERIAL NO: 120029

BASE: Roughdeck 4ft x 4ft

SERIAL NO: 59377795PU

CAPACITY: 2500 kgx0.5 kg (5000lbs x 1lbs)

CORNER TEST:



Front

Total Test Weight Applied

1000 lbs

Corner	Applied Weight	<u>As Found</u>		<u>As Left</u>		
		Scale Reading	Scale Error	Applied Weight	Scale Reading	Scale Error
1	1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs
2	1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs
3	1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs
4	1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs
5	1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs

<u>Applied Weight</u>	<u>As Found</u>	<u>Scale Error</u>	<u>Applied Weight</u>	<u>As Left</u>	<u>Scale Error</u>
0 lbs	0.0 lbs	0.0 lbs	0 lbs	0.0 lbs	0.0 lbs
1000 lbs	1000.5 lbs	0.5 lbs	1000 lbs	1000.0 lbs	0.0 lbs
2000 lbs	2001.5 lbs	1.5 lbs	2000 lbs	2000.0 lbs	0.0 lbs
3000 lbs	3002.0 lbs	2.0 lbs	3000 lbs	3000.0 lbs	0.0 lbs
4000 lbs	4002.5 lbs	2.5 lbs	4000 lbs	4000.0 lbs	0.0 lbs
5000 lbs	5005.5 lbs	5.5 lbs	5000 lbs	5000.0 lbs	0.0 lbs
0 lbs	0.0 lbs	0.0 lbs	0 lbs	0.0 lbs	0.0 lbs

Weights Used: 5x1000lbs S/N BS1, BS2, BS3, BS4, BS6.

Comments: Recalibrated and tested - OK

Note: Next quarterly inspection due April 2008